

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120424\
 Data File : VN085085.D
 Acq On : 04 Dec 2024 11:27
 Operator : JC\MD
 Sample : VN1204MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1204MBL01

Quant Time: Dec 04 22:01:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	159453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	274668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	231111	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	92393	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	115818	50.992	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.980%
35) Dibromofluoromethane	8.165	113	93712	50.593	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.180%
50) Toluene-d8	10.565	98	315629	47.548	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.100%
62) 4-Bromofluorobenzene	12.847	95	104026	41.905	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	83.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120424\
Data File : VN085085.D
Acq On : 04 Dec 2024 11:27
Operator : JC\MD
Sample : VN1204MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1204MBL01

Quant Time: Dec 04 22:01:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

