

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120424\
 Data File : VN085089.D
 Acq On : 04 Dec 2024 13:13
 Operator : JC\MD
 Sample : PB165342TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB165342TB

Quant Time: Dec 04 22:05:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	152627	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	266172	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	236674	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	93039	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	109020	50.146	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 100.300%			
35) Dibromofluoromethane	8.165	113	88439	49.270	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 98.540%			
50) Toluene-d8	10.565	98	320546	49.830	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.660%			
62) 4-Bromofluorobenzene	12.847	95	106506	44.273	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 88.540%			
Target Compounds						
16) Acetone	4.436	43	36459	57.582	ug/l	Qvalue 100
43) Isopropyl Acetate	8.683	43	126080	27.462	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120424\
Data File : VN085089.D
Acq On : 04 Dec 2024 13:13
Operator : JC\MD
Sample : PB165342TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB165342TB

Quant Time: Dec 04 22:05:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration

