

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120424\
 Data File : VN085086.D
 Acq On : 04 Dec 2024 11:50
 Operator : JC\MD
 Sample : VN1204WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1204WBL01

Quant Time: Dec 04 22:02:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 23 02:54:10 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	155240	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	271880	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	231756	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	89427	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	113375	51.271	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	102.540%	
35) Dibromofluoromethane	8.165	113	93450	50.969	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	101.940%	
50) Toluene-d8	10.565	98	310609	47.272	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	94.540%	
62) 4-Bromofluorobenzene	12.847	95	106097	43.177	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	86.360%	


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120424\
Data File : VN085086.D
Acq On : 04 Dec 2024 11:50
Operator : JC\MD
Sample : VN1204WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1204WBL01

Quant Time: Dec 04 22:02:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112224W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 23 02:54:10 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.224 min Scan# 1066
Delta R.T. 0.006 min
Lab File: VN085086.D
Acq: 04 Dec 2024 11:50




Instrument :
MSVOA_N
ClientSampleId :
VN1204WBL01


Tgt Ion:168 Resp: 155240
Ion Ratio Lower Upper
168 100
99 61.3 51.2 76.8



#33
1,2-Dichloroethane-d4
Concen: 51.271 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. 0.006 min
Lab File: VN085086.D
Acq: 04 Dec 2024 11:50


Tgt Ion: 65 Resp: 113375
Ion Ratio Lower Upper
65 100
67 51.3 0.0 105.0








#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 9.100 min Scan# 1215
 Delta R.T. 0.006 min
 Lab File: VN085086.D
 Acq: 04 Dec 2024 11:50

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1204WBL01







Tgt Ion:	114	Resp:	271880
Ion	Ratio	Lower	Upper
114	100		
63	21.3	0.0	41.0
88	16.2	0.0	34.8



#35
 Dibromofluoromethane
 Concen: 50.969 ug/l
 RT: 8.165 min Scan# 1056
 Delta R.T. 0.006 min
 Lab File: VN085086.D
 Acq: 04 Dec 2024 11:50

Tgt Ion:	113	Resp:	93450
Ion	Ratio	Lower	Upper
113	100		
111	102.3	82.0	123.0
192	18.9	14.8	22.2





#50

Toluene-d8

Concen: 47.272 ug/l

RT: 10.565 min Scan# 1464

Delta R.T. -0.000 min

Lab File: VN085086.D

Acq: 04 Dec 2024 11:50

Instrument :

MSVOA_N

ClientSampleId :





VN1204WBL01

Tgt Ion: 98 Resp: 310609

Ion Ratio Lower Upper

98 100

100 66.4 52.5 78.7



#62

4-Bromofluorobenzene

Concen: 43.177 ug/l

RT: 12.847 min Scan# 1852

Delta R.T. -0.000 min

Lab File: VN085086.D

Acq: 04 Dec 2024 11:50




Tgt Ion: 95 Resp: 106097


Ion Ratio Lower Upper

95 100

174 74.3 0.0 146.6

176 71.8 0.0 141.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.865 min Scan# 1684

Delta R.T. 0.006 min

Lab File: VN085086.D

Acq: 04 Dec 2024 11:50

Instrument :

MSVOA_N

ClientSampleId :

VN1204WBL01





Tgt Ion:117 Resp: 231756

Ion Ratio Lower Upper

117 100

82 57.6 46.0 69.0

119 31.8 25.6 38.4



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.788 min Scan# 2012

Delta R.T. -0.000 min

Lab File: VN085086.D

Acq: 04 Dec 2024 11:50

Tgt Ion:152 Resp: 89427

Ion Ratio Lower Upper

152 100

115 62.5 31.7 95.1

150 160.7 0.0 351.4

