

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088438.D
 Acq On : 04 Dec 2025 12:17
 Operator : JC\MD
 Sample : VN1204MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1204MBL01

Quant Time: Dec 05 00:12:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	158186	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	365050	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	366228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	166667	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	197070	66.133	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	132.260%#
35) Dibromofluoromethane	8.147	113	145632	57.567	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	115.140%
50) Toluene-d8	10.541	98	494184	52.502	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.000%
62) 4-Bromofluorobenzene	12.829	95	177477	54.952	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	109.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088438.D
Acq On : 04 Dec 2025 12:17
Operator : JC\MD
Sample : VN1204MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1204MBL01

Quant Time: Dec 05 00:12:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

