

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088445.D
 Acq On : 04 Dec 2025 15:06
 Operator : JC\MD
 Sample : Q3756-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BR-01-190-120225

Quant Time: Dec 05 00:16:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	246275	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	564051	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	564375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	278056	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	260534	56.158	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.320%
35) Dibromofluoromethane	8.147	113	179677	45.967	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.940%
50) Toluene-d8	10.541	98	640125	44.013	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.020%
62) 4-Bromofluorobenzene	12.829	95	263053	52.713	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.420%
Target Compounds						
4) Vinyl Chloride	2.542	62	102480	25.356	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088445.D
Acq On : 04 Dec 2025 15:06
Operator : JC\MD
Sample : Q3756-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BR-01-190-120225

Quant Time: Dec 05 00:16:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

