

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088448.D
 Acq On : 04 Dec 2025 16:07
 Operator : JC\MD
 Sample : Q3756-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-13B-47.5-120225

Quant Time: Dec 05 00:17:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	206071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	471902	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	469221	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	222578	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	216155	55.682	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.360%
35) Dibromofluoromethane	8.147	113	154264	47.172	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.340%
50) Toluene-d8	10.547	98	550181	45.216	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.440%
62) 4-Bromofluorobenzene	12.829	95	211464	50.650	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088448.D
Acq On : 04 Dec 2025 16:07
Operator : JC\MD
Sample : Q3756-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-13B-47.5-120225

Quant Time: Dec 05 00:17:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

