

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088451.D
 Acq On : 04 Dec 2025 17:09
 Operator : JC\MD
 Sample : Q3757-01 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-14A-17.5-120325

Quant Time: Dec 05 00:18:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	203888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	466381	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	458828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	216202	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	210028	54.683	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.360%	
35) Dibromofluoromethane	8.153	113	146713	45.394	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 90.780%	
50) Toluene-d8	10.541	98	532452	44.277	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 88.560%	
62) 4-Bromofluorobenzene	12.829	95	199907	48.448	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 96.900%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088451.D
Acq On : 04 Dec 2025 17:09
Operator : JC\MD
Sample : Q3757-01 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-14A-17.5-120325

Quant Time: Dec 05 00:18:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

