

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088452.D
 Acq On : 04 Dec 2025 17:29
 Operator : JC\MD
 Sample : Q3757-03 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BR-03D-490-120325

Quant Time: Dec 05 00:18:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	204906	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.082	114	477681	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	464655	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	216573	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	217079	56.238	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.480%
35) Dibromofluoromethane	8.147	113	154647	46.717	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.440%
50) Toluene-d8	10.541	98	550915	44.728	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.460%
62) 4-Bromofluorobenzene	12.829	95	208569	49.352	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	98.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088452.D
Acq On : 04 Dec 2025 17:29
Operator : JC\MD
Sample : Q3757-03 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BR-03D-490-120325

Quant Time: Dec 05 00:18:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

