

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088461.D
 Acq On : 04 Dec 2025 20:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 05 00:22:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	103	0.00
2 T	Dichlorodifluoromethane	0.751	0.718	4.4	100	0.00
3 P	Chloromethane	0.807	0.732	9.3	94	0.00
4 C	Vinyl Chloride	0.821	0.780	5.0#	99	0.00
5 T	Bromomethane	0.395	0.356	9.9	92	0.00
6 T	Chloroethane	0.508	0.493	3.0	100	0.00
7 T	Trichlorofluoromethane	1.153	1.145	0.7	108	0.00
8 T	Diethyl Ether	0.433	0.456	-5.3	106	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.614	0.590	3.9	108	0.00
10 T	Methyl Iodide	0.608	0.510	16.1	79	0.00
11 T	Tert butyl alcohol	0.150	0.136	9.3	87	0.00
12 CM	1,1-Dichloroethene	0.610	0.589	3.4#	107	0.00
13 T	Acrolein	0.144	0.172	-19.4	117	0.00
14 T	Allyl chloride	1.131	1.146	-1.3	103	0.00
15 T	Acrylonitrile	0.434	0.436	-0.5	96	0.00
16 T	Acetone	0.341	0.427	-25.2#	131	0.00
17 T	Carbon Disulfide	1.993	1.818	8.8	94	0.00
18 T	Methyl Acetate	1.014	1.107	-9.2	109	0.00
19 T	Methyl tert-butyl Ether	2.285	2.481	-8.6	102	0.00
20 T	Methylene Chloride	0.729	0.730	-0.1	99	0.00
21 T	trans-1,2-Dichloroethene	0.676	0.673	0.4	101	0.00
22 T	Diisopropyl ether	2.414	2.668	-10.5	106	0.00
23 T	Vinyl Acetate	1.945	2.354	-21.0	113	0.00
24 P	1,1-Dichloroethane	1.367	1.432	-4.8	103	0.00
25 T	2-Butanone	0.548	0.619	-13.0	107	0.00
26 T	2,2-Dichloropropane	1.180	1.130	4.2	97	0.00
27 T	cis-1,2-Dichloroethene	0.785	0.805	-2.5	100	0.00
28 T	Bromochloromethane	0.681	0.739	-8.5	104	0.00
29 T	Tetrahydrofuran	0.393	0.416	-5.9	100	0.00
30 C	Chloroform	1.343	1.449	-7.9#	107	0.00
31 T	Cyclohexane	1.239	1.118	9.8	102	0.00
32 T	1,1,1-Trichloroethane	1.181	1.226	-3.8	109	0.00
33 S	1,2-Dichloroethane-d4	0.942	1.104	-17.2	114	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.346	0.380	-9.8	112	0.00
36 T	1,1-Dichloropropene	0.535	0.507	5.2	106	0.00
37 T	Ethyl Acetate	0.703	0.689	2.0	98	0.00
38 T	Carbon Tetrachloride	0.532	0.535	-0.6	110	0.00
39 T	Methylcyclohexane	0.572	0.502	12.2	98	0.00
40 TM	Benzene	1.570	1.538	2.0	101	0.00
41 T	Methacrylonitrile	0.348	0.359	-3.2	103	0.00
42 TM	1,2-Dichloroethane	0.599	0.644	-7.5	108	0.00
43 T	Isopropyl Acetate	1.074	1.105	-2.9	102	0.00
44 TM	Trichloroethene	0.371	0.331	10.8	95	0.00
45 C	1,2-Dichloropropane	0.402	0.408	-1.5#	103	0.00
46 T	Dibromomethane	0.287	0.297	-3.5	104	0.00
47 T	Bromodichloromethane	0.586	0.612	-4.4	107	0.00
48 T	Methyl methacrylate	0.490	0.525	-7.1	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088461.D
 Acq On : 04 Dec 2025 20:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 05 00:22:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.005	16.7	78	0.00
50 S	Toluene-d8	1.289	1.390	-7.8	111	0.00
51 T	4-Methyl-2-Pentanone	0.633	0.674	-6.5	103	0.00
52 CM	Toluene	0.906	0.935	-3.2#	105	0.00
53 T	t-1,3-Dichloropropene	0.601	0.620	-3.2	101	0.00
54 T	cis-1,3-Dichloropropene	0.632	0.653	-3.3	101	0.00
55 T	1,1,2-Trichloroethane	0.351	0.361	-2.8	105	0.00
56 T	Ethyl methacrylate	0.554	0.624	-12.6	104	0.00
57 T	1,3-Dichloropropane	0.633	0.660	-4.3	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.319	0.357	-11.9	101	0.00
59 T	2-Hexanone	0.399	0.452	-13.3	107	0.00
60 T	Dibromochloromethane	0.400	0.419	-4.7	105	0.00
61 T	1,2-Dibromoethane	0.355	0.363	-2.3	101	0.00
62 S	4-Bromofluorobenzene	0.442	0.514	-16.3	118	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	107	0.00
64 T	Tetrachloroethene	0.396	0.317	19.9	90	0.00
65 PM	Chlorobenzene	1.123	1.083	3.6	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.373	1.6	104	0.00
67 C	Ethyl Benzene	1.963	1.933	1.5#	107	0.00
68 T	m/p-Xylenes	0.728	0.717	1.5	105	0.00
69 T	o-Xylene	0.694	0.681	1.9	105	0.00
70 T	Styrene	1.127	1.207	-7.1	108	0.00
71 P	Bromoform	0.277	0.287	-3.6	108	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
73 T	Isopropylbenzene	3.697	3.492	5.5	106	0.00
74 T	N-amyl acetate	1.259	1.168	7.2	123	0.00
75 P	1,1,2,2-Tetrachloroethane	1.195	1.159	3.0	113	0.00
76 T	1,2,3-Trichloropropane	1.171	1.101	6.0	101	0.00
77 T	Bromobenzene	0.846	0.816	3.5	104	0.00
78 T	n-propylbenzene	4.523	4.251	6.0	106	0.00
79 T	2-Chlorotoluene	2.762	2.650	4.1	108	0.00
80 T	1,3,5-Trimethylbenzene	3.063	2.971	3.0	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.413	0.408	1.2	105	0.00
82 T	4-Chlorotoluene	2.743	2.736	0.3	106	0.00
83 T	tert-Butylbenzene	2.604	2.442	6.2	107	0.00
84 T	1,2,4-Trimethylbenzene	3.045	3.033	0.4	107	0.00
85 T	sec-Butylbenzene	3.741	3.444	7.9	105	0.00
86 T	p-Isopropyltoluene	3.068	2.884	6.0	104	0.00
87 T	1,3-Dichlorobenzene	1.647	1.543	6.3	106	0.00
88 T	1,4-Dichlorobenzene	1.725	1.588	7.9	104	0.00
89 T	n-Butylbenzene	2.979	2.664	10.6	102	0.00
90 T	Hexachloroethane	0.560	0.518	7.5	107	0.00
91 T	1,2-Dichlorobenzene	1.550	1.485	4.2	108	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.290	0.273	5.9	104	0.00
93 T	1,2,4-Trichlorobenzene	0.921	0.810	12.1	100	0.00
94 T	Hexachlorobutadiene	0.331	0.259	21.8	96	0.00
95 T	Naphthalene	3.190	3.005	5.8	101	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088461.D
Acq On : 04 Dec 2025 20:33
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 05 00:22:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.896	0.769	14.2	97	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6