

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
 Data File : VN088454.D
 Acq On : 04 Dec 2025 18:10
 Operator : JC\MD
 Sample : Q3757-07 100X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RMW-05B-90.5-120325

Quant Time: Dec 05 00:19:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 25 01:34:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.206	168	215953	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.083	114	502668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.841	117	496953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.770	152	236823	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.559	65	234326	57.601	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.200%
35) Dibromofluoromethane	8.147	113	166231	47.720	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.440%
50) Toluene-d8	10.547	98	583555	45.023	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	90.040%
62) 4-Bromofluorobenzene	12.829	95	225173	50.632	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	101.260%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120425\
Data File : VN088454.D
Acq On : 04 Dec 2025 18:10
Operator : JC\MD
Sample : Q3757-07 100X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RMW-05B-90.5-120325

Quant Time: Dec 05 00:19:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 25 01:34:52 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.206 min Scan# 1063
Delta R.T. 0.000 min
Lab File: VN088454.D
Acq: 04 Dec 2025 18:10




Instrument :
MSVOA_N
ClientSampleId :
RMW-05B-90.5-120325


Tgt Ion:168 Resp: 215953
Ion Ratio Lower Upper
168 100
99 75.7 57.1 85.7



#33
1,2-Dichloroethane-d4
Concen: 57.601 ug/l
RT: 8.559 min Scan# 1123
Delta R.T. 0.000 min
Lab File: VN088454.D
Acq: 04 Dec 2025 18:10

Tgt Ion: 65 Resp: 234326
Ion Ratio Lower Upper
65 100
67 50.3 0.0 100.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.083 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN088454.D

Acq: 04 Dec 2025 18:10

Instrument :

MSVOA_N

ClientSampleId :

RMW-05B-90.5-120325





Tgt Ion:114 Resp: 502668

Ion Ratio Lower Upper

114 100

63 23.2 0.0 49.0

88 16.9 0.0 34.0



#35

Dibromofluoromethane

Concen: 47.720 ug/l

RT: 8.147 min Scan# 1053

Delta R.T. -0.006 min

Lab File: VN088454.D

Acq: 04 Dec 2025 18:10




Tgt Ion:113 Resp: 166231


Ion Ratio Lower Upper

113 100

111 102.7 82.5 123.7

192 15.6 12.8 19.2









#50
Toluene-d8
Concen: 45.023 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN088454.D
Acq: 04 Dec 2025 18:10




Instrument :
MSVOA_N
ClientSampleId :
RMW-05B-90.5-120325


Tgt Ion: 98 Resp: 583555
Ion Ratio Lower Upper
98 100
100 64.4 52.2 78.2



#62
4-Bromofluorobenzene
Concen: 50.632 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.006 min
Lab File: VN088454.D
Acq: 04 Dec 2025 18:10

Tgt Ion: 95 Resp: 225173
Ion Ratio Lower Upper
95 100
174 67.1 0.0 139.0
176 63.1 0.0 132.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.841 min Scan# 1
Delta R.T. -0.006 min
Lab File: VN088454.D
Acq: 04 Dec 2025 18:10

Instrument :
MSVOA_N
ClientSampleId :
RMW-05B-90.5-120325

Tgt Ion:117 Resp: 496953
Ion Ratio Lower Upper
117 100
82 63.1 47.8 71.8
119 31.9 26.9 40.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.770 min Scan# 2009
Delta R.T. 0.000 min
Lab File: VN088454.D
Acq: 04 Dec 2025 18:10

Tgt Ion:152 Resp: 236823
Ion Ratio Lower Upper
152 100
115 66.8 33.0 98.9
150 157.7 0.0 349.0

