

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120518\
 Data File : VN052670.D
 Acq On : 5 Dec 2018 12:26
 Operator : MD\SY
 Sample : PB115330ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115330ZHE#01

Quant Time: Dec 06 04:52:59 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1174157	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1809905	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1579502	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	557221	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	619660	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
35) Dibromofluoromethane	7.59	113	561935	50.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.72%	
50) Toluene-d8	10.09	98	2219828	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
62) 4-Bromofluorobenzene	12.40	95	671367	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
Target Compounds						
16) Acetone	3.82	43	18759	5.563	ug/l #	87
20) Methylene Chloride	4.55	84	31308	2.373	ug/l	96
43) Isopropyl Acetate	8.17	43	47356	2.683	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120518\
Data File : VN052670.D
Acq On : 5 Dec 2018 12:26
Operator : MD\SY
Sample : PB115330ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115330ZHE#01

Quant Time: Dec 06 04:52:59 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

