

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120518\
 Data File : VN052672.D
 Acq On : 5 Dec 2018 13:20
 Operator : MD\SY
 Sample : PB115330ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115330ZHE#03

Quant Time: Dec 06 04:58:19 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1149012	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1790916	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1587646	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	587057	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	608756	48.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.32%	
35) Dibromofluoromethane	7.59	113	553249	50.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.22%	
50) Toluene-d8	10.09	98	2184829	50.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.68%	
62) 4-Bromofluorobenzene	12.40	95	686238	47.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.04%	
Target Compounds						
16) Acetone	3.82	43	18454	5.592	ug/l	97
20) Methylene Chloride	4.55	84	30664	2.375	ug/l	97
43) Isopropyl Acetate	8.17	43	54683	3.131	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120518\
Data File : VN052672.D
Acq On : 5 Dec 2018 13:20
Operator : MD\SY
Sample : PB115330ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115330ZHE#03

Quant Time: Dec 06 04:58:19 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

