

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120518\
 Data File : VN052674.D
 Acq On : 5 Dec 2018 14:13
 Operator : MD\SY
 Sample : PB115330ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115330ZHE#05

Quant Time: Dec 06 05:01:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1226886	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1913076	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1712939	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	647809	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	655051	49.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.08%	
35) Dibromofluoromethane	7.59	113	586683	49.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.48%	
50) Toluene-d8	10.09	98	2344981	51.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.16%	
62) 4-Bromofluorobenzene	12.40	95	750635	48.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.32%	
Target Compounds						
16) Acetone	3.82	43	16114	4.573	ug/l	97
20) Methylene Chloride	4.55	84	85993	6.237	ug/l	97
43) Isopropyl Acetate	8.17	43	61164	3.278	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120518\
Data File : VN052674.D
Acq On : 5 Dec 2018 14:13
Operator : MD\SY
Sample : PB115330ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115330ZHE#05

Quant Time: Dec 06 05:01:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

