

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120518\
 Data File : VN052676.D
 Acq On : 5 Dec 2018 15:06
 Operator : MD\SY
 Sample : PB115330ZHE#07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 13 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115330ZHE#07

Quant Time: Dec 06 05:06:16 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1123839	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1762410	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1546701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	572817	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	597829	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	7.59	113	546374	50.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.58%	
50) Toluene-d8	10.09	98	2149086	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	12.40	95	666267	46.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.76%	
Target Compounds						
16) Acetone	3.82	43	17230	5.338	ug/l	95
20) Methylene Chloride	4.55	84	30720	2.432	ug/l	96
43) Isopropyl Acetate	8.17	43	54495	3.170	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120518\
Data File : VN052676.D
Acq On : 5 Dec 2018 15:06
Operator : MD\SY
Sample : PB115330ZHE#07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 13 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115330ZHE#07

Quant Time: Dec 06 05:06:16 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

