

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120518\
 Data File : VN052680.D
 Acq On : 5 Dec 2018 16:53
 Operator : MD\SY
 Sample : PB115330ZHE#11
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115330ZHE#11

Quant Time: Dec 06 05:45:45 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1155904	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1800314	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1591039	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	586562	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	622417	49.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.90%	
35) Dibromofluoromethane	7.59	113	556931	50.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.36%	
50) Toluene-d8	10.09	98	2216568	51.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.62%	
62) 4-Bromofluorobenzene	12.40	95	682084	46.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.96%	
Target Compounds						
16) Acetone	3.82	43	17963	5.411	ug/l	96
20) Methylene Chloride	4.55	84	31975	2.461	ug/l	91
43) Isopropyl Acetate	8.17	43	73858	4.206	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120518\
Data File : VN052680.D
Acq On : 5 Dec 2018 16:53
Operator : MD\SY
Sample : PB115330ZHE#11
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115330ZHE#11

Quant Time: Dec 06 05:45:45 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

