

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120518\
 Data File : VN052682.D
 Acq On : 5 Dec 2018 17:47
 Operator : MD\SY
 Sample : J6202-02RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SAMPLE-1RE

Quant Time: Dec 06 05:54:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1189893	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1832776	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1568092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	548507	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	626721	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
35) Dibromofluoromethane	7.59	113	319454	28.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	56.54%	
50) Toluene-d8	10.09	98	2227538	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
62) 4-Bromofluorobenzene	12.40	95	659909	44.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.30%	
Target Compounds						
16) Acetone	3.82	43	161492	47.258	ug/l	99
20) Methylene Chloride	4.55	84	364363	27.247	ug/l	98
95) Naphthalene	15.14	128	17193	4.204	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120518\
Data File : VN052682.D
Acq On : 5 Dec 2018 17:47
Operator : MD\SY
Sample : J6202-02RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SAMPLE-1RE

Quant Time: Dec 06 05:54:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

