

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120519\
 Data File : VN059492.D
 Acq On : 5 Dec 2019 12:02
 Operator : JC/SP
 Sample : VN1205WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBL01

Quant Time: Dec 06 00:41:47 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	332804	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	513373	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	437379	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	148063	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	201746	48.80	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.60%	
35) Dibromofluoromethane	7.57	113	156727	49.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.04%	
50) Toluene-d8	10.08	98	626005	49.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.32%	
62) 4-Bromofluorobenzene	12.40	95	187281	42.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.08%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120519\
Data File : VN059492.D
Acq On : 5 Dec 2019 12:02
Operator : JC/SP
Sample : VN1205WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBL01

Quant Time: Dec 06 00:41:47 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

