

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120519\
 Data File : VN059494.D
 Acq On : 5 Dec 2019 12:58
 Operator : JC/SP
 Sample : PB125248TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB125248TB

Quant Time: Dec 06 00:46:30 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	324379	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	513352	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	463418	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	176389	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	204654	50.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.58%	
35) Dibromofluoromethane	7.57	113	159138	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
50) Toluene-d8	10.08	98	634081	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.40	95	207554	47.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.28%	
Target Compounds						
43) Isopropyl Acetate	8.15	43	100470	11.320	ug/l #	91
95) Naphthalene	15.13	128	11943	3.558	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120519\
Data File : VN059494.D
Acq On : 5 Dec 2019 12:58
Operator : JC/SP
Sample : PB125248TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB125248TB

Quant Time: Dec 06 00:46:30 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

