

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120519\
 Data File : VN059495.D
 Acq On : 5 Dec 2019 13:20
 Operator : JC/SP
 Sample : K6112-04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-120419

Quant Time: Dec 06 00:51:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	311040	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	490608	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	435123	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	156414	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	197952	51.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.46%	
35) Dibromofluoromethane	7.57	113	151239	49.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.00%	
50) Toluene-d8	10.08	98	606035	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
62) 4-Bromofluorobenzene	12.40	95	190931	45.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.76%	
Target Compounds						
20) Methylene Chloride	4.52	84	2709687	761.594	ug/l	99
43) Isopropyl Acetate	8.15	43	76266	8.992	ug/l #	90
95) Naphthalene	15.13	128	9848	3.479	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120519\
Data File : VN059495.D
Acq On : 5 Dec 2019 13:20
Operator : JC/SP
Sample : K6112-04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-120419

Quant Time: Dec 06 00:51:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

