

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120519\
 Data File : VN059498.D
 Acq On : 5 Dec 2019 14:26
 Operator : JC/SP
 Sample : K6135-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 COMP-36

Quant Time: Dec 06 01:00:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	307554	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	485264	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	435574	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	159515	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	199484	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
35) Dibromofluoromethane	7.57	113	150809	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
50) Toluene-d8	10.08	98	608551	51.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.14%	
62) 4-Bromofluorobenzene	12.40	95	190142	45.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.38%	
Target Compounds						
20) Methylene Chloride	4.52	84	8193	1.104	ug/l	93
43) Isopropyl Acetate	8.14	43	84623	10.087	ug/l #	89
95) Naphthalene	15.13	128	10304	3.506	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120519\
Data File : VN059498.D
Acq On : 5 Dec 2019 14:26
Operator : JC/SP
Sample : K6135-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
COMP-36

Quant Time: Dec 06 01:00:41 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

