

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120523\
 Data File : VN080359.D
 Acq On : 05 Dec 2023 12:25
 Operator : JC\MD
 Sample : 05612-07 1000X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-06I-DUP

Quant Time: Dec 06 00:49:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	153611	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	285463	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	304122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	137523	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	141131	55.643	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 111.280%	
35) Dibromofluoromethane	8.165	113	110393	52.410	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.820%	
50) Toluene-d8	10.565	98	377565	49.726	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.460%	
62) 4-Bromofluorobenzene	12.853	95	148133	51.651	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.300%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120523\
Data File : VN080359.D
Acq On : 05 Dec 2023 12:25
Operator : JC\MD
Sample : 05612-07 1000X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-06I-DUP

Quant Time: Dec 06 00:49:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

