

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120523\
 Data File : VN080364.D
 Acq On : 05 Dec 2023 14:23
 Operator : JC\MD
 Sample : 05622-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-5

Quant Time: Dec 06 00:51:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	164546	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	291311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	302981	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	132096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	143584	52.848	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 105.700%	
35) Dibromofluoromethane	8.165	113	115148	53.570	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.140%	
50) Toluene-d8	10.570	98	391304	50.501	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 101.000%	
62) 4-Bromofluorobenzene	12.853	95	146609	50.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.180%	
Target Compounds						
					Qvalue	
30) Chloroform	7.965	83	4514	1.145	ug/l	86
64) Tetrachloroethene	11.112	164	4449	1.924	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120523\
Data File : VN080364.D
Acq On : 05 Dec 2023 14:23
Operator : JC\MD
Sample : 05622-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-5

Quant Time: Dec 06 00:51:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

