

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120523\
 Data File : VN080373.D
 Acq On : 05 Dec 2023 18:08
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 06 00:55:20 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	176372	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	332880	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	353039	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	161921	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	158881	54.557	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.120%
35) Dibromofluoromethane	8.171	113	125906	51.260	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.520%
50) Toluene-d8	10.570	98	433343	48.942	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.853	95	177400	53.044	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.080%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120523\
Data File : VN080373.D
Acq On : 05 Dec 2023 18:08
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 06 00:55:20 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

