

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120525\
 Data File : VN088466.D
 Acq On : 05 Dec 2025 10:26
 Operator : JC\MD
 Sample : VN1205WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1205WBL01

Quant Time: Dec 05 23:38:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111325W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 14 00:49:21 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.794	128	52610	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	265051	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.841	117	235330	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	158711	33.026	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	110.100%#
60) 4-Bromofluorobenzene	12.829	95	105075	26.413	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	88.033%
63) Toluene-d8	10.541	98	331315	29.737	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	99.133%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120525\
Data File : VN088466.D
Acq On : 05 Dec 2025 10:26
Operator : JC\MD
Sample : VN1205WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1205WBL01

Quant Time: Dec 05 23:38:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111325W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 14 00:49:21 2025
Response via : Initial Calibration

