

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120618\
 Data File : VN052691.D
 Acq On : 6 Dec 2018 11:04
 Operator : MD\SY
 Sample : PB115376TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115376TB

Quant Time: Dec 07 06:56:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1157179	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1807542	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1565715	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	574969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	626702	49.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.48%	
35) Dibromofluoromethane	7.59	113	420867	37.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	75.54%	
50) Toluene-d8	10.09	98	2200025	50.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.44%	
62) 4-Bromofluorobenzene	12.40	95	678046	46.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.04%	
Target Compounds						
16) Acetone	3.82	43	122516	36.866	ug/l	98
20) Methylene Chloride	4.55	84	246255	18.936	ug/l	97
95) Naphthalene	15.14	128	11102	3.906	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120618\
Data File : VN052691.D
Acq On : 6 Dec 2018 11:04
Operator : MD\SY
Sample : PB115376TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115376TB

Quant Time: Dec 07 06:56:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

