

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120618\
 Data File : VN052695.D
 Acq On : 6 Dec 2018 13:13
 Operator : MD\SY
 Sample : PB115376ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115376ZHE#04

Quant Time: Dec 07 07:02:36 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	934724	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.58	114	1478971	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1335144	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	519158	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.02	65	510210	50.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.26%	
35) Dibromofluoromethane	7.59	113	266419	29.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	58.44%	
50) Toluene-d8	10.09	98	1829520	51.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.10%	
62) 4-Bromofluorobenzene	12.40	95	595761	49.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.90%	
Target Compounds						
16) Acetone	3.81	43	139329	51.903	ug/l	100
20) Methylene Chloride	4.54	84	293858	27.974	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120618\
Data File : VN052695.D
Acq On : 6 Dec 2018 13:13
Operator : MD\SY
Sample : PB115376ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115376ZHE#04

Quant Time: Dec 07 07:02:36 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

