

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120618\
 Data File : VN052697.D
 Acq On : 6 Dec 2018 14:07
 Operator : MD\SY
 Sample : PB115376ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB115376ZHE#06

Quant Time: Dec 07 07:05:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1132586	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1767241	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1547701	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	558083	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	608685	49.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.72%	
35) Dibromofluoromethane	7.59	113	329236	30.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	60.44%	
50) Toluene-d8	10.09	98	2144535	50.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.14%	
62) 4-Bromofluorobenzene	12.40	95	653177	45.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.66%	
Target Compounds						
16) Acetone	3.82	43	155490	47.804	ug/l	99
20) Methylene Chloride	4.55	84	318764	25.043	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120618\
Data File : VN052697.D
Acq On : 6 Dec 2018 14:07
Operator : MD\SY
Sample : PB115376ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB115376ZHE#06

Quant Time: Dec 07 07:05:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

