

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120618\
 Data File : VN052700.D
 Acq On : 6 Dec 2018 15:27
 Operator : MD\SY
 Sample : J6214-37
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 TW-10

Quant Time: Dec 07 07:11:53 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
 Quant Title : SW846 8260
 QLast Update : Mon Dec 03 03:24:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	1080555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1681381	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1426542	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	472373	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	574800	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
35) Dibromofluoromethane	7.59	113	523299	50.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.96%	
50) Toluene-d8	10.09	98	2035604	50.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.90%	
62) 4-Bromofluorobenzene	12.40	95	600205	44.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.54%	
Target Compounds						
16) Acetone	3.83	43	13116	4.227	ug/l	90
95) Naphthalene	15.13	128	6769	3.783	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120618\
Data File : VN052700.D
Acq On : 6 Dec 2018 15:27
Operator : MD\SY
Sample : J6214-37
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
TW-10

Quant Time: Dec 07 07:11:53 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112718W.M
Quant Title : SW846 8260
QLast Update : Mon Dec 03 03:24:02 2018
Response via : Initial Calibration

