

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120619\
 Data File : VN059506.D
 Acq On : 6 Dec 2019 11:22
 Operator : JC/SP
 Sample : VN1206WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1206WBL01

Quant Time: Dec 06 22:38:26 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	230848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	368990	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	329152	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	119371	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	164218	57.26	ug/l	0.00
Spiked Amount	50.000		Recovery	=	114.52%	
35) Dibromofluoromethane	7.57	113	119153	51.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.70%	
50) Toluene-d8	10.08	98	462833	51.08	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.16%	
62) 4-Bromofluorobenzene	12.40	95	141422	44.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.38%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120619\
Data File : VN059506.D
Acq On : 6 Dec 2019 11:22
Operator : JC/SP
Sample : VN1206WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1206WBL01

Quant Time: Dec 06 22:38:26 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

