

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120619\
 Data File : VN059510.D
 Acq On : 6 Dec 2019 13:25
 Operator : JC/SP
 Sample : K6162-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC

Quant Time: Dec 06 22:49:44 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	240313	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	384677	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	344782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	119066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	171594	57.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.96%	
35) Dibromofluoromethane	7.57	113	123648	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
50) Toluene-d8	10.08	98	486311	51.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.96%	
62) 4-Bromofluorobenzene	12.40	95	148133	44.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.80%	
Target Compounds						
20) Methylene Chloride	4.52	84	221799	79.589	ug/l	95
43) Isopropyl Acetate	8.15	43	65795	9.893	ug/l #	89
95) Naphthalene	15.13	128	5090	3.141	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120619\
Data File : VN059510.D
Acq On : 6 Dec 2019 13:25
Operator : JC/SP
Sample : K6162-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC

Quant Time: Dec 06 22:49:44 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

