

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120619\
 Data File : VN059511.D
 Acq On : 6 Dec 2019 13:47
 Operator : JC/SP
 Sample : K6147-12
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-WC001

Quant Time: Dec 06 22:57:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	229069	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	367755	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	334165	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	123403	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	165276	58.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.16%	
35) Dibromofluoromethane	7.57	113	117673	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
50) Toluene-d8	10.08	98	465973	51.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.20%	
62) 4-Bromofluorobenzene	12.40	95	145083	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
Target Compounds						
16) Acetone	3.80	43	28584	13.054	ug/l	93
43) Isopropyl Acetate	8.15	43	51912	8.165	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120619\
Data File : VN059511.D
Acq On : 6 Dec 2019 13:47
Operator : JC/SP
Sample : K6147-12
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-WC001

Quant Time: Dec 06 22:57:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

