

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120619\
 Data File : VN059521.D
 Acq On : 6 Dec 2019 17:27
 Operator : JC/SP
 Sample : K6147-22
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P001-WC011

Quant Time: Dec 06 23:33:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
 Quant Title : SW846 8260
 QLast Update : Tue Dec 03 16:37:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	228165	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	368014	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	331454	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	118875	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	167321	59.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.06%	
35) Dibromofluoromethane	7.57	113	119428	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
50) Toluene-d8	10.08	98	463416	51.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.56%	
62) 4-Bromofluorobenzene	12.40	95	143343	45.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.84%	
Target Compounds						
16) Acetone	3.79	43	34010	16.628	ug/l	93
18) Methyl Acetate	4.30	43	8266	2.166	ug/l #	90
43) Isopropyl Acetate	8.15	43	52023	8.177	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120619\
Data File : VN059521.D
Acq On : 6 Dec 2019 17:27
Operator : JC/SP
Sample : K6147-22
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P001-WC011

Quant Time: Dec 06 23:33:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N120319W.M
Quant Title : SW846 8260
QLast Update : Tue Dec 03 16:37:42 2019
Response via : Initial Calibration

