

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
 Data File : VN069782.D
 Acq On : 6 Dec 2021 14:35
 Operator : JC/MD
 Sample : M4893-01
 Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EO-1-120121

Quant Time: Dec 06 15:14:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.080	168	1393514	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.965	114	2464844	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2368632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	1088069	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.432	65	1068680	49.669	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.340%
35) Dibromofluoromethane	8.018	113	709466	45.952	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	91.900%
50) Toluene-d8	10.440	98	3070050	49.522	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	99.040%
62) 4-Bromofluorobenzene	12.733	95	1177924	52.448	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
Data File : VN069782.D
Acq On : 6 Dec 2021 14:35
Operator : JC/MD
Sample : M4893-01
Misc : 5.04g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EO-1-120121

Quant Time: Dec 06 15:14:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

