

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
 Data File : VN069788.D
 Acq On : 6 Dec 2021 17:10
 Operator : JC/MD
 Sample : M4922-12DL 5X
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DUP02-20211130DL

Quant Time: Dec 07 03:49:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1335638	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2331170	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2419661	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	935667	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	971176	47.094	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.180%
35) Dibromofluoromethane	8.024	113	675537	46.263	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.520%
50) Toluene-d8	10.443	98	2955372	50.406	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.820%
62) 4-Bromofluorobenzene	12.734	95	1142155	53.771	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.540%
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.210	101	32407	2.455	ug/l	98
12) 1,1-Dichloroethene	4.170	96	13749	1.081	ug/l	92
30) Chloroform	7.823	83	9537	0.325	ug/l	99
44) Trichloroethene	9.217	130	558808	34.515	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
Data File : VN069788.D
Acq On : 6 Dec 2021 17:10
Operator : JC/MD
Sample : M4922-12DL 5X
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DUP02-20211130DL

Quant Time: Dec 07 03:49:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

