

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
 Data File : VN069792.D
 Acq On : 6 Dec 2021 18:53
 Operator : JC/MD
 Sample : M4950-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Dec 07 03:50:07 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1280832	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2257044	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2394115	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	902508	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	946785	47.875	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	95.760%
35) Dibromofluoromethane	8.024	113	664425	46.997	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	94.000%
50) Toluene-d8	10.443	98	2914201	51.336	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.680%
62) 4-Bromofluorobenzene	12.733	95	1101650	53.568	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.140%
Target Compounds						
16) Acetone	4.293	43	16657	1.400	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
Data File : VN069792.D
Acq On : 6 Dec 2021 18:53
Operator : JC/MD
Sample : M4950-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANAG

Quant Time: Dec 07 03:50:07 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

