

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
 Data File : VN069793.D
 Acq On : 6 Dec 2021 19:19
 Operator : JC/MD
 Sample : PB141160TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB141160TB

Quant Time: Dec 07 03:50:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1363918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	2443021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	2614536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	1025693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	1016432	48.266	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.540%
35) Dibromofluoromethane	8.024	113	706911	46.196	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	92.400%
50) Toluene-d8	10.443	98	3143870	51.166	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.340%
62) 4-Bromofluorobenzene	12.733	95	1226465	55.097	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	110.200%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	155130	12.246	ug/l	95
20) Methylene Chloride	5.095	84	25682	1.525	ug/l	90
43) Isopropyl Acetate	8.566	43	158446	3.093	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
Data File : VN069793.D
Acq On : 6 Dec 2021 19:19
Operator : JC/MD
Sample : PB141160TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB141160TB

Quant Time: Dec 07 03:50:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

