

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
 Data File : VN069777.D
 Acq On : 6 Dec 2021 11:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 06 15:11:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	50.000	52.830	-5.7	94	0.00
3 P	Chloromethane	50.000	48.218	3.6	92	0.00
4 C	Vinyl Chloride	50.000	50.356	-0.7#	94	0.00
5 T	Bromomethane	50.000	50.588	-1.2	94	0.00
6 T	Chloroethane	50.000	49.063	1.9	93	0.00
7 T	Trichlorofluoromethane	50.000	50.980	-2.0	96	0.00
8 T	Diethyl Ether	50.000	50.358	-0.7	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.488	-7.0	101	0.00
10 T	Methyl Iodide	50.000	44.708	10.6	80	0.00
11 T	Tert butyl alcohol	250.000	217.937	12.8	81	-0.01
12 CM	1,1-Dichloroethene	50.000	50.188	-0.4#	95	0.00
13 T	Acrolein	250.000	375.138	-50.1#	150	0.00
14 T	Allyl chloride	50.000	50.466	-0.9	95	0.00
15 T	Acrylonitrile	250.000	234.002	6.4	86	0.00
16 T	Acetone	250.000	259.450	-3.8	96	0.00
17 T	Carbon Disulfide	50.000	52.705	-5.4	97	0.00
18 T	Methyl Acetate	50.000	45.757	8.5	87	0.00
19 T	Methyl tert-butyl Ether	50.000	49.971	0.1	93	0.00
20 T	Methylene Chloride	50.000	48.264	3.5	96	0.00
21 T	trans-1,2-Dichloroethene	50.000	50.401	-0.8	95	0.00
22 T	Diisopropyl ether	50.000	51.765	-3.5	95	0.00
23 T	Vinyl Acetate	250.000	260.484	-4.2	92	0.00
24 P	1,1-Dichloroethane	50.000	51.717	-3.4	97	0.00
25 T	2-Butanone	250.000	236.788	5.3	86	0.00
26 T	2,2-Dichloropropane	50.000	53.422	-6.8	101	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.089	-0.2	93	0.00
28 T	Bromochloromethane	50.000	48.787	2.4	92	0.00
29 T	Tetrahydrofuran	250.000	234.213	6.3	84	0.00
30 C	Chloroform	50.000	50.953	-1.9#	95	0.00
31 T	Cyclohexane	50.000	50.225	-0.5	96	0.00
32 T	1,1,1-Trichloroethane	50.000	51.513	-3.0	94	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.788	4.4	91	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	97	0.00
35 S	Dibromofluoromethane	50.000	50.007	-0.0	90	0.00
36 T	1,1-Dichloropropene	50.000	53.786	-7.6	95	0.00
37 T	Ethyl Acetate	50.000	47.421	5.2	84	0.00
38 T	Carbon Tetrachloride	50.000	54.317	-8.6	96	0.00
39 T	Methylcyclohexane	50.000	56.475	-13.0	98	0.00
40 TM	Benzene	50.000	52.691	-5.4	94	0.00
41 T	Methacrylonitrile	50.000	50.696	-1.4	87	0.00
42 TM	1,2-Dichloroethane	50.000	51.693	-3.4	93	0.00
43 T	Isopropyl Acetate	50.000	50.731	-1.5	89	0.00
44 TM	Trichloroethene	50.000	53.054	-6.1	94	0.00
45 C	1,2-Dichloropropane	50.000	52.859	-5.7#	95	0.00
46 T	Dibromomethane	50.000	52.516	-5.0	93	0.00
47 T	Bromodichloromethane	50.000	55.461	-10.9	96	0.00
48 T	Methyl methacrylate	50.000	51.251	-2.5	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
 Data File : VN069777.D
 Acq On : 6 Dec 2021 11:30
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 06 15:11:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 03 08:46:47 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	982.308	1.8	86	0.00
50 S	Toluene-d8	50.000	49.712	0.6	89	0.00
51 T	4-Methyl-2-Pentanone	250.000	254.823	-1.9	86	0.00
52 CM	Toluene	50.000	53.908	-7.8#	94	0.00
53 T	t-1,3-Dichloropropene	50.000	49.468	1.1	94	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.984	-2.0	95	0.00
55 T	1,1,2-Trichloroethane	50.000	52.169	-4.3	92	0.00
56 T	Ethyl methacrylate	50.000	47.334	5.3	89	0.00
57 T	1,3-Dichloropropane	50.000	52.734	-5.5	93	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	213.848	14.5	84	0.00
59 T	2-Hexanone	250.000	253.273	-1.3	86	0.00
60 T	Dibromochloromethane	50.000	50.048	-0.1	95	0.00
61 T	1,2-Dibromoethane	50.000	52.752	-5.5	92	0.00
62 S	4-Bromofluorobenzene	50.000	51.442	-2.9	93	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	97	0.00
64 T	Tetrachloroethene	50.000	52.893	-5.8	95	0.00
65 PM	Chlorobenzene	50.000	52.767	-5.5	94	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	54.642	-9.3	94	0.00
67 C	Ethyl Benzene	50.000	55.175	-10.3#	95	0.00
68 T	m/p-Xylenes	100.000	111.863	-11.9	96	0.00
69 T	o-Xylene	50.000	55.072	-10.1	95	0.00
70 T	Styrene	50.000	56.518	-13.0	94	0.00
71 P	Bromoform	50.000	49.402	1.2	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	99	0.00
73 T	Isopropylbenzene	50.000	50.465	-0.9	96	0.00
74 T	N-amyl acetate	50.000	48.937	2.1	89	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	47.168	5.7	91	0.00
76 T	1,2,3-Trichloropropane	50.000	51.237	-2.5	97	0.00
77 T	Bromobenzene	50.000	49.006	2.0	94	0.00
78 T	n-propylbenzene	50.000	52.847	-5.7	97	0.00
79 T	2-Chlorotoluene	50.000	50.328	-0.7	96	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.907	-3.8	97	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	48.065	3.9	92	0.00
82 T	4-Chlorotoluene	50.000	51.974	-3.9	98	0.00
83 T	tert-Butylbenzene	50.000	50.096	-0.2	94	0.00
84 T	1,2,4-Trimethylbenzene	50.000	52.907	-5.8	96	0.00
85 T	sec-Butylbenzene	50.000	52.724	-5.4	96	0.00
86 T	p-Isopropyltoluene	50.000	54.177	-8.4	98	0.00
87 T	1,3-Dichlorobenzene	50.000	52.403	-4.8	98	0.00
88 T	1,4-Dichlorobenzene	50.000	51.167	-2.3	98	0.00
89 T	n-Butylbenzene	50.000	55.585	-11.2	99	0.00
90 T	Hexachloroethane	50.000	53.195	-6.4	95	0.00
91 T	1,2-Dichlorobenzene	50.000	50.891	-1.8	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	47.572	4.9	87	0.00
93 T	1,2,4-Trichlorobenzene	50.000	48.070	3.9	95	0.00
94 T	Hexachlorobutadiene	50.000	53.205	-6.4	98	0.00
95 T	Naphthalene	50.000	41.016	18.0	83	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120621\
Data File : VN069777.D
Acq On : 6 Dec 2021 11:30
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Dec 06 15:11:18 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N120221W.M
Quant Title : SW846 8260
QLast Update : Fri Dec 03 08:46:47 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	47.374	5.3	92	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6