

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080377.D
 Acq On : 06 Dec 2023 11:27
 Operator : JC\MD
 Sample : VN1206MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1206MBL01

Quant Time: Dec 07 00:55:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	157388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	287749	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	305041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	130399	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	137150	52.776	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	105.560%
35) Dibromofluoromethane	8.165	113	110627	52.104	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.200%
50) Toluene-d8	10.571	98	378134	49.405	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.820%
62) 4-Bromofluorobenzene	12.853	95	146070	50.527	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080377.D
Acq On : 06 Dec 2023 11:27
Operator : JC\MD
Sample : VN1206MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1206MBL01

Quant Time: Dec 07 00:55:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

