

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080378.D
 Acq On : 06 Dec 2023 11:50
 Operator : JC\MD
 Sample : VN1206WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1206WBL01

Quant Time: Dec 07 00:56:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	174815	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	315153	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	337917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145190	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	150397	52.104	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.200%
35) Dibromofluoromethane	8.165	113	121344	52.182	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.360%
50) Toluene-d8	10.571	98	419137	50.001	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.000%
62) 4-Bromofluorobenzene	12.853	95	157339	49.692	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080378.D
Acq On : 06 Dec 2023 11:50
Operator : JC\MD
Sample : VN1206WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1206WBL01

Quant Time: Dec 07 00:56:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

