

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080382.D
 Acq On : 06 Dec 2023 13:36
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 07 00:59:05 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	214368	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	373118	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	389186	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	189040	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	164739	46.542	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.080%
35) Dibromofluoromethane	8.165	113	137626	49.989	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.980%
50) Toluene-d8	10.571	98	495791	49.957	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.920%
62) 4-Bromofluorobenzene	12.853	95	205122	54.719	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080382.D
Acq On : 06 Dec 2023 13:36
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 07 00:59:05 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

