

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080387.D
 Acq On : 06 Dec 2023 15:35
 Operator : JC\MD
 Sample : PB157577TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB157577TB

Quant Time: Dec 07 01:01:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	185620	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	342717	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	374579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	163424	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	159927	52.181	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.360%
35) Dibromofluoromethane	8.171	113	126046	49.844	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.680%
50) Toluene-d8	10.565	98	467470	51.281	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.560%
62) 4-Bromofluorobenzene	12.847	95	186304	54.108	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.220%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	17663	17.158	ug/l #	82
20) Methylene Chloride	5.289	84	15358	7.124	ug/l #	91
25) 2-Butanone	7.489	43	16974	13.878	ug/l	95
43) Isopropyl Acetate	8.688	43	100085	20.252	ug/l #	87
95) Naphthalene	15.641	128	23056	2.476	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080387.D
Acq On : 06 Dec 2023 15:35
Operator : JC\MD
Sample : PB157577TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB157577TB

Quant Time: Dec 07 01:01:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

