

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080393.D
 Acq On : 06 Dec 2023 17:57
 Operator : JC\MD
 Sample : 05631-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2A

Quant Time: Dec 07 01:04:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	168166	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	308739	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	324612	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	145346	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	150423	54.174	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.340%
35) Dibromofluoromethane	8.165	113	118324	51.940	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.880%
50) Toluene-d8	10.571	98	411560	50.117	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.240%
62) 4-Bromofluorobenzene	12.853	95	162330	52.334	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080393.D
Acq On : 06 Dec 2023 17:57
Operator : JC\MD
Sample : 05631-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2A

Quant Time: Dec 07 01:04:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

