

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080397.D
 Acq On : 06 Dec 2023 19:31
 Operator : JC\MD
 Sample : 05650-04 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 50197

Quant Time: Dec 07 01:06:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	162149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	292448	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	320341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	166421	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	144417	53.941	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 107.880%	
35) Dibromofluoromethane	8.171	113	117255	54.338	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.680%	
50) Toluene-d8	10.570	98	398706	51.256	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.520%	
62) 4-Bromofluorobenzene	12.853	95	161167	54.853	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 109.700%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	16536	18.389	ug/l	95
25) 2-Butanone	7.494	43	5920	5.541	ug/l	88
95) Naphthalene	15.647	128	6096	0.643	ug/l #	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080397.D
Acq On : 06 Dec 2023 19:31
Operator : JC\MD
Sample : 05650-04 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
50197

Quant Time: Dec 07 01:06:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

