

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
 Data File : VN080399.D
 Acq On : 06 Dec 2023 20:19
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 07 01:07:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 17 01:19:32 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	206092	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	359417	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	379457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	191450	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	161784	47.543	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 95.080%			
35) Dibromofluoromethane	8.165	113	136329	51.406	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 102.820%			
50) Toluene-d8	10.571	98	477783	49.977	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 99.960%			
62) 4-Bromofluorobenzene	12.847	95	209311	57.965	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 115.940%			
Target Compounds						
95) Naphthalene	15.641	128	355688	32.612	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120623\
Data File : VN080399.D
Acq On : 06 Dec 2023 20:19
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 07 01:07:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N111623W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 17 01:19:32 2023
Response via : Initial Calibration

