


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN120623\  
Data File : VN080375.D  
Acq On : 06 Dec 2023 08:48  
Operator : JC\MD  
Sample : BFB  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N111623W.M  
Title : SW846 8260  
Last Update : Fri Nov 17 01:19:32 2023



AutoFind: Scans 1851, 1852, 1853; Background Corrected with Scan 1844

| Target | Rel. to | Lower  | Upper  | Rel.  | Raw   | Result    |
|--------|---------|--------|--------|-------|-------|-----------|
| Mass   | Mass    | Limit% | Limit% | Abn%  | Abn   | Pass/Fail |
| 50     | 95      | 15     | 40     | 15.8  | 6800  | PASS      |
| 75     | 95      | 30     | 60     | 55.5  | 23957 | PASS      |
| 95     | 95      | 100    | 100    | 100.0 | 43144 | PASS      |
| 96     | 95      | 5      | 9      | 6.6   | 2859  | PASS      |
| 173    | 174     | 0.00   | 2      | 1.7   | 563   | PASS      |
| 174    | 95      | 50     | 100    | 79.0  | 34077 | PASS      |
| 175    | 174     | 5      | 9      | 6.5   | 2210  | PASS      |
| 176    | 174     | 95     | 101    | 98.1  | 33427 | PASS      |
| 177    | 176     | 5      | 9      | 6.6   | 2219  | PASS      |

Average of 12.841 to 12.853 min.: VN080375.D\data.ms

BFB

Modified:subtracted

| m/z   | abund. | m/z   | abund. | m/z   | abund. | m/z   | abund. |
|-------|--------|-------|--------|-------|--------|-------|--------|
| 36.00 | 355    | 49.05 | 1537   | 63.00 | 1939   | 72.95 | 2051   |
| 37.05 | 1644   | 50.05 | 6800   | 63.90 | 127    | 74.05 | 7851   |
| 38.05 | 1544   | 51.00 | 2130   | 64.05 | 288    | 75.05 | 2395   |
| 38.90 | 634    | 52.10 | 200    | 65.10 | 547    | 76.00 | 1971   |
| 39.20 | 178    | 55.80 | 172    | 67.10 | 60     | 77.05 | 200    |
| 40.00 | 77     | 56.00 | 542    | 68.00 | 4336   | 77.90 | 126    |
| 44.05 | 363    | 56.90 | 306    | 69.05 | 4482   | 78.95 | 1794   |
| 45.00 | 466    | 57.10 | 847    | 69.90 | 132    | 79.90 | 491    |
| 46.85 | 213    | 60.00 | 449    | 70.15 | 292    | 80.85 | 2063   |
| 47.15 | 273    | 61.05 | 2233   | 71.85 | 183    | 81.95 | 315    |
| 48.15 | 233    | 62.00 | 2346   | 72.10 | 103    | 86.95 | 1266   |

Average of 12.841 to 12.853 min.: VN080375.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z    | abund. | m/z    | abund. | m/z    | abund. |
|--------|--------|--------|--------|--------|--------|--------|--------|
| 88.00  | 1545   | 105.90 | 124    | 129.70 | 96     | 147.95 | 141    |
| 90.70  | 66     | 106.10 | 54     | 130.00 | 83     | 160.70 | 70     |
| 91.10  | 52     | 115.70 | 108    | 133.10 | 55     | 171.10 | 51     |
| 91.90  | 416    | 116.00 | 104    | 137.00 | 61     | 171.60 | 83     |
| 92.05  | 1013   | 116.80 | 96     | 140.70 | 291    | 171.90 | 404    |
| 93.05  | 2303   | 116.95 | 312    | 140.95 | 451    | 172.20 | 97     |
| 94.05  | 5492   | 117.60 | 51     | 142.85 | 903    | 172.95 | 563    |
| 95.00  | 43144  | 117.85 | 166    | 144.70 | 229    | 173.95 | 34077  |
| 96.00  | 2859   | 118.70 | 56     | 144.90 | 150    | 175.00 | 2210   |
| 103.85 | 240    | 119.00 | 383    | 145.75 | 192    | 175.95 | 33427  |
| 105.75 | 154    | 127.90 | 195    | 146.90 | 72     | 177.00 | 2219   |

Average of 12.841 to 12.853 min.: VN080375.D\data.ms

BFB

Modified:subtracted

| m/z    | abund. | m/z | abund. | m/z | abund. | m/z | abund. |
|--------|--------|-----|--------|-----|--------|-----|--------|
| 192.90 | 51     |     |        |     |        |     |        |
| 266.75 | 124    |     |        |     |        |     |        |
| 280.80 | 195    |     |        |     |        |     |        |
| 281.25 | 464    |     |        |     |        |     |        |
| 281.95 | 200    |     |        |     |        |     |        |
| 283.00 | 84     |     |        |     |        |     |        |
| 355.10 | 56     |     |        |     |        |     |        |