

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120624\
 Data File : VN085122.D
 Acq On : 06 Dec 2024 12:01
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Dec 06 23:05:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 06 22:50:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	30628	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	158053	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	142637	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.576	65	72572	29.862	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	99.533%
60) 4-Bromofluorobenzene	12.847	95	58934	24.971	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	83.233%
63) Toluene-d8	10.565	98	192599	28.785	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	95.967%
Target Compounds						
16) Carbon Disulfide	4.712	76	10841	1.944	ug/l	# 93
90) Hexachlorobutadiene	15.500	225	4525	4.984	ug/l	96
91) Naphthalene	15.635	128	19465	3.586	ug/l	98
92) 1,2,3-Trichlorobenzene	15.835	180	7668	4.389	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120624\
Data File : VN085122.D
Acq On : 06 Dec 2024 12:01
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Dec 06 23:05:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 06 22:50:42 2024
Response via : Initial Calibration

