

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120624\
 Data File : VN085132.D
 Acq On : 06 Dec 2024 17:15
 Operator : JC\MD
 Sample : P5068-02 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 14B-2

Quant Time: Dec 06 23:11:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Dec 06 22:50:42 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.812	128	28577	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.100	114	131902	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.865	117	120856	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.577	65	67283	29.673	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.900%
60) 4-Bromofluorobenzene	12.847	95	48276	24.142	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	80.467%
63) Toluene-d8	10.565	98	162898	28.734	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	95.767%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN120624\
Data File : VN085132.D
Acq On : 06 Dec 2024 17:15
Operator : JC\MD
Sample : P5068-02 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
14B-2

Quant Time: Dec 06 23:11:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N120624W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Dec 06 22:50:42 2024
Response via : Initial Calibration

