

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120720\
 Data File : VN064947.D
 Acq On : 7 Dec 2020 16:00
 Operator : JC/MD
 Sample : L4954-19
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SS013-13-6711-120220

Quant Time: Dec 08 16:57:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	181279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	318708	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	338464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	159568	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	106583	50.57	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.14%	
35) Dibromofluoromethane		7.55	113	98947	51.88	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.76%	
50) Toluene-d8		10.06	98	402502	51.66	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.32%	
62) 4-Bromofluorobenzene		12.37	95	162674	58.33	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	116.66%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	3.78	43	30812	44.475	ug/l	89
39) Methylcyclohexane	9.04	83	29202	9.183	ug/l	95
67) Ethyl Benzene	11.48	91	175189	14.428	ug/l	100
68) m/p-Xylenes	11.59	106	80237	17.453	ug/l	98
69) o-Xylene	11.92	106	33670	7.718	ug/l	95
73) Isopropylbenzene	12.22	105	224988	20.271	ug/l	100
78) n-propylbenzene	12.57	91	183281	14.495	ug/l	100
80) 1,3,5-Trimethylbenzene	12.71	105	52019	5.639	ug/l	99
83) tert-Butylbenzene	12.96	119	3447	0.446	ug/l	93
84) 1,2,4-Trimethylbenzene	13.01	105	330534	35.436	ug/l	100
85) sec-Butylbenzene	13.14	105	27423	2.671	ug/l	79
86) p-Isopropyltoluene	13.26	119	10341	1.106	ug/l	98
95) Naphthalene	15.10	128	158388	16.372	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120720\
Data File : VN064947.D
Acq On : 7 Dec 2020 16:00
Operator : JC/MD
Sample : L4954-19
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SS013-13-6711-120220

Quant Time: Dec 08 16:57:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

