

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120720\
 Data File : VN064951.D
 Acq On : 7 Dec 2020 17:36
 Operator : JC/MD
 Sample : L4962-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-4

Quant Time: Dec 08 16:55:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.62	168	176378	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	312171	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	322946	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	143523	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		7.99	65	107222	52.29	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.58%	
35) Dibromofluoromethane		7.55	113	98127	52.53	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.06%	
50) Toluene-d8		10.06	98	386787	50.68	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.36%	
62) 4-Bromofluorobenzene		12.38	95	150200	54.99	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	109.98%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120720\
Data File : VN064951.D
Acq On : 7 Dec 2020 17:36
Operator : JC/MD
Sample : L4962-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-4

Quant Time: Dec 08 16:55:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

