

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN120720\
 Data File : VN064952.D
 Acq On : 7 Dec 2020 18:00
 Operator : JC/MD
 Sample : L4962-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Quant Time: Dec 08 08:34:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 23 13:54:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	175018	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	307179	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	317276	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	140780	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	105308	51.76	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery	=	103.52%	
35) Dibromofluoromethane	7.55	113	99374	54.06	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery	=	108.12%	
50) Toluene-d8	10.06	98	381691	50.82	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery	=	101.64%	
62) 4-Bromofluorobenzene	12.38	95	148939	55.41	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery	=	110.82%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN120720\
Data File : VN064952.D
Acq On : 7 Dec 2020 18:00
Operator : JC/MD
Sample : L4962-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Quant Time: Dec 08 08:34:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N112320W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 23 13:54:12 2020
Response via : Initial Calibration

